

उत्तर प्रदेश पुलिस तकनीकी सेवायें मुख्यालय,

चतुर्थ तल, जवाहर भवन, अशोक मार्ग, लखनऊ
संख्या-टीएस-एमओडी-120/2015-16(एफएसएल) दिनांक -जनवरी, 08 2016


निविदा सूचना

- 1- उत्तर प्रदेश पुलिस हेतु निम्न उपकरणों की सील बन्द निविदा भारतीय मैनुफैक्चरर/डीलर/रजिस्टर्ड फर्म /विदेशी मैनुफैक्चरर के अधिकृत विक्रेता से आमंत्रित की जाती है।
- 2- निविदा प्रपत्र व उपकरण से सम्बन्धित तकनीकी स्पेशीफिकेशन उत्तर प्रदेश पुलिस की वेबसाइट <http://uppolice.gov.in> से डाउनलोड किया जा सकता है। निविदा शुल्क के रूप में ₹0 300.00 का बैंक ड्राफ्ट/बैंकर चेक/एफडीआर पुलिस अधीक्षक/सहायक निदेशक, उ0प्र0पुलिस कम्प्यूटर केन्द्र लखनऊ के नाम देय होगा, जमा की जायेगी।
- 3- निविदा शुल्क किसी भी दशा में वापस नहीं किया जायेगा तथा निविदा स्वीकृत न होने की दशा में ईएमडी वापस कर दी जायेगी।
- 4- तकनीकी सेवायें मुख्यालय से नगद भुगतान पर टेण्डर डाकुमेंट्स विक्रय नहीं किये जायेंगे।
- 5- प्रत्येक टेण्डरकर्ता द्वारा टेण्डर की ईएमडीए डीडी/ एफडीआर / बैंक गारन्टी के रूप में, जो पुलिस अधीक्षक/सहायक निदेशक, उ0प्र0पुलिस कम्प्यूटर केन्द्र लखनऊ के नाम देय होगा, जमा की जायेगी।
- 6- सफल टेण्डरकर्ता द्वारा कयादेश मूल्य का 10 प्रतिशत धनराशि सिक्क्योरिटी मनी के रूप में जमा किया जायेगा, जो उपकरण के सन्तोषजनक ढंग से कार्य करने के उपरान्त उनके वारन्टी अवधि के उपरान्त वापस कर दी जायेगी।
- 7- तकनीकी स्पेशीफिकेशन के अनुसार प्रोप्राइटरी आइटम या उनके समतुल्य उत्पाद/सेवा अन्तराष्ट्रीय स्तर के स्वीकार होंगे। प्रोप्राइटरी उत्पाद या इसके समतुल्य उत्पाद/सेवा सम्बन्धित तकनीकी समिति का निर्णय अन्तिम निर्णय होगा।
- 8- उपकरणों की संख्या विभाग की आवश्यकतानुसार बढ़ायी जा सकती है तथा विभाग द्वारा बढ़ायी गयी संख्या को फर्म को निर्धारित अवधि में आपूर्ति करना होगा।
- 9- निम्न किसी उपकरण के तकनीकी स्पेशीफिकेशन में यदि कोई संशोधन किया जाता है तो निविदा जमा करने की अन्तिम तिथि के 07 दिवस पूर्व ही संशोधित तकनीकी स्पेशीफिकेशन अपलोड कर दिया जायेगा।
- 10- समस्त टैक्स सहित न्यूनतम निविदा एल-1 पर विचारा किया जायेगा।
- 11- तकनीकी मुख्यालय द्वारा जनहित में किसी भी समय बिना कारण दर्शाये टेण्डर को निरस्त किया जा सकता है।

टेण्डर संख्या	आइटम	सं०	ईएमडी (₹0 में)	टेण्डर डाकुमेंट का मूल्य (₹0 में)
1-टीएस-मोड-130/15(एफएसएल)	रमन स्पेक्ट्रोमीटर	01	1,70,000-00	300.00
2-टीएस-मोड-132/15(एफएसएल)	एक्स0आर0एफ0 स्पेक्ट्रोमीटर विथ एसेसिरीज	01	1,30,000-00	300.00

निविदा जमा करने की अन्तिम तिथि
तकनीकी भाव पत्र खोले जाने की तिथि
तकनीकी भाव पत्र पर कय समिति की बैठक
वित्तीय भाव पत्र खोले जाने की तिथि
वित्तीय भाव पत्र पर कय समिति की बैठक

22.01.2016 समय 02.00 बजे अपरान्ह
22.01.2016 समय 04.00 बजे अपरान्ह
25.01.2016 समय 11.30 बजे पूर्वान्ह
27.01.2016 समय 11.30 बजे पूर्वान्ह
28.11.2016 समय 04.00 बजे अपरान्ह



(संजय कुमार)

अपर पुलिस अधीक्षक, तकनीकी सेवायें,
निमित्त- पुलिस महानिरीक्षक, तकनीकी सेवायें,
लखनऊ।

Raman Spectrometer

Specifications:

1. Laser wavelength 532nm, 638/633nm, 785nm with laser output power 200mW.
2. Laser with filters.
3. Automatic switching between white light and laser.
4. Auto calibration.
5. Protection with laser facility.
6. Raman spectral library.
7. Spectral resolution 1 cm^{-1} or better.
8. Spectral range $100\text{-}3000\text{ cm}^{-1}$ (wave numbers).
9. Analytical software package for examination of Forensic exhibits.
10. Provision to examine exhibits like ink, Paints, Passport, Laminates, Fibres etc.
11. Research grade trinocular microscope with objectives 5x, 20x, 40x/50x and reflected light illumination.
12. Sampling accessories for solid and powdered sample.
13. CCD array detector, peltier cooling.
14. High resolution colour camera (5MP or more).
15. High end PC with license softwares with accessories having 32" TFT monitor.
16. Colour laser printer of high speed and high resolution.
17. Online UPS 2KVA having 2Hrs backup.
18. Warranty two years onsite.
19. AMC & service : 5 Year.
20. License and software upgradation.
21. Training : free of cost for Scientists.

XRF SPECTROMETER WITH ACCESSORIES

Technical Specifications:

- Instrument should have light weight handheld XRF analyzer
- Integrated touch-screen color display, ...
- Fundamental and Empirical calibration (*both)
- Fast start-up: No standardization needed
- It should have calibration facility
- It should have 40 / 50 kV Power / Fixed filter for alloy application OR X-ray tube for accurate result.
- Easy data transfer with USB Flash drive
- Sharp Beam® technology - Optimized detector and tube geometry
Provides excellent light element,, performance
- Customize calibration can at sight without the help of PC
- Pen Drive and USB Cable for data transfer
- USB port to get result/spectrum directly from machine at site OR Data can be transfer via Bluetooth / USB by laptop / PC.
- Better accuracy and repeatability in SDD model compare to SI PIN detector model .
- Spare and consumable available in Local currency in India for future service
- Continuous analysis in Field without cooling or any other problem.
- SDD detector in basic model (without LE facility) .
- Better Beam Path
- Smart Library
- IP54 Certified
- 2 year comprehensive warranty
- More than 8 Hours individual battery life for a single full charge
- One High end laptop with Pre. Installed softwares & with latest Windows O.S.-8 / 10.
- Installation and Installation room preparation and workstation with air conditioning.